

# MONTHLY PROBE THERMOMETER CHECK

Probe thermometer recording details

Month												
	Reading in Iced Water											
	Reading in Boiling Water											
	Checked by											

- The readings in **iced water should be -1°C to +1°C**, if outside this range the unit should be replaced or returned to the manufacturer to be recalibrated.
- The reading in **boiling water should be between 99°C and 101°C**, if outside this range the unit should be replaced or returned to the manufacturer to be recalibrated.

Date and Details of Yearly Calibration		Date of next Yearly Calibration	
--	--	---------------------------------	--

**N.B. The electronic display unit should be checked at least once per year. Manufacturers may offer a calibration service.**